

NEO-SERIES

compatible to: *NEOSS Implant System® / NEOSS™*

NEW



PLATFORM

3,5 - 5,5 mm

TITANIUM BASE

for individual milled
Zirconium Abutment
incl. screw
Titanium Grade 5 ELI



NEO 80355500

SCAN BODY 3D-GUIDE®

for intra + extra
oral scanning




NEO 9.S3D3.555



NT-MILL LAB ANALOG

Stainless Steel



NEO 5.La3.555



IMPLANT PICK-UP

for open-tray with
integrated screw
Coming soon



NEO TR-00013.5

ABUTMENT SCREW

Recommended tightening
torques 30 Ncm
Titanium Grade 5 ELI



NEO 603555

Notice: Products indicated with ® and ™ are registered brand names of the respective manufactures.
Products indicated with ★ are available in the USA 

NEXT GENERATION DESIGN – MADE IN GERMANY ///

NEW

NEO-SERIES

compatible to: NEOSS Implant System® / NEOSS™



PLATFORM

3,5 - 5,5 mm

2-CONNECT®-ABUTMENT SET

incl. 2-CONnect®-Coping
and Coping-screw
Titanium Grade 5 ELI



Gingiva Height	0,5 mm	NEO-2CON 8035551S
	2,5 mm	NEO-2CON 8035552S

2-CONNECT®-ABUTMENT

Recommended tightening
torques: 30 Ncm
Titanium Grade 5 ELI

M (Male Part)



Gingiva Height	0,5 mm	NEO-2CON 8035551M
	2,5 mm	NEO-2CON 8035552M

2-CONNECT®-COPING

Titanium Grade 5 ELI



F (Female Part)

2CON F-U



COPING-SCREW

Recommended tightening
torques: 25 Ncm
Titanium Grade 5 ELI



KS 61



NEO-SERIES

compatible to: NEOSS Implant System® / NEOSS™

NEW



PLATFORM

3,5 - 5,5 mm

2-CONNECT®-ABUTMENT SET

incl. 2-CONnect®-Coping,
Coping-screw & Male Part Screw
angled 20°

Gingiva Height 2,5 mm

Titanium Grade 5 ELI

Coming soon



NEO-2CON 80355A3S

2-CONNECT®-ABUTMENT

angled 20°

Gingiva Height 2,5 mm

Titanium Grade 5 ELI

Coming soon

M (Male Part)



NEO-2CON 80355A3M

2-CONNECT®-COPING

Titanium Grade 5 ELI



F (Female Part)

2CON F-U



COPING-SCREW

Recommended tightening
torques: 15 Ncm

Titanium Grade 5 ELI

Coming soon



KS 62

MALE PART SCREW

for angled 20° 2-CONnect®

Abutment | Recommended tightening
torques: 20 Ncm

Titanium Grade 5 ELI

Coming soon



NEO 6.Ms3.555

Notice: Products indicated with ® and ™ are registered brand names of the respective manufactures.
Products indicated with ★ are available in the USA 

NEXT GENERATION DESIGN – MADE IN GERMANY ///

NEO-SERIES

compatible to: NEOSS Implant System® / NEOSS™



PLATFORM

3,5 - 5,5 mm

2-CONNECT®- IMPLANT PICK-UP

integrated screw



● Straight KS 61 = ø 1,8 mm	2CON TR-00021.8
● Angled KS 62 = ø 1,6 mm	2CON TR-00031.6

2-CONNECT®- SCAN BODY 3D-GUIDE®

for intra + extra
oral scanning



● Straight KS 61 = ø 1,8 mm	2CON 9.S3D1.800	★
● Angled KS 62 = ø 1,6 mm	2CON 9.S3D1.600	★

2-CONNECT®- LAB ANALOG



● Straight KS 61 = ø 1,8 mm	2CON 5.La1.800	★
● Angled KS 62 = ø 1,6 mm	2CON 5.La1.600	★

2-CONNECT®- DIM ANALOG



● Straight KS 61 = ø 1,8 mm	2CON 5.DIM.180	★
● Angled KS 62 = ø 1,6 mm	2CON 5.DIM.160	★

SLEEVE

for plaster model
Stainless Steel



DIM 5.XSL.A30	★
---------------	---

NEO-SERIES

compatible to: NEOSS Implant System® / NEOSS™

NEW



PLATFORM

3,5 - 5,5 mm

NT-PREFORM®

incl. coloured milling screw
Titanium Grade 5 ELI




Gingiva Height 0,5 mm

NEO 10.Mi3.555.05

NT-MILL LAB ANALOG

for nt-Preform®
Stainless Steel



NEO 5.La3.555



ABUTMENT GRIP

for nt-Holder fixture
Stainless Steel



W 12.200.000



SCAN BODY 3D-GUIDE®

for intra + extra
oral scanning



NEO 9.S3D3.555



COLOURED MILLING SCREW

for nt-Preform®
Recommended tightening
torques 10 Ncm



NEO 6.Sc3.555



ABUTMENT SCREW

Recommended
tightening torques 30 Ncm
Titanium Grade 5 ELI



NEO 603555

Notice: Products indicated with ® and ™ are registered brand names of the respective manufactures.
Products indicated with ★ are available in the USA 

NEXT GENERATION DESIGN – MADE IN GERMANY ///

NEW

NEO-SERIES DIM ANALOG

compatible to: *NEOSS Implant System® / NEOSS™*



PLATFORM

3,5 - 5,5 mm

DIM ANALOG



NEO 5.DIM.355



SLEEVE

for plaster model
Stainless Steel



DIM 5.XSL.A20

